

INSTRUCTION MANUAL



CS215 Temperature and Relative Humidity Probe

Revision: 4/16



Copyright © 2005-2016
Campbell Scientific, Inc.

Limited Warranty

“Products manufactured by CSI are warranted by CSI to be free from defects in materials and workmanship under normal use and service for twelve months from the date of shipment unless otherwise specified in the corresponding product manual. (Product manuals are available for review online at www.campbellsci.com.) Products not manufactured by CSI, but that are resold by CSI, are warranted only to the limits extended by the original manufacturer. Batteries, fine-wire thermocouples, desiccant, and other consumables have no warranty. CSI’s obligation under this warranty is limited to repairing or replacing (at CSI’s option) defective Products, which shall be the sole and exclusive remedy under this warranty. The Customer assumes all costs of removing, reinstalling, and shipping defective Products to CSI. CSI will return such Products by surface carrier prepaid within the continental United States of America. To all other locations, CSI will return such Products best way CIP (port of entry) per Incoterms ® 2010. This warranty shall not apply to any Products which have been subjected to modification, misuse, neglect, improper service, accidents of nature, or shipping damage. This warranty is in lieu of all other warranties, expressed or implied. The warranty for installation services performed by CSI such as programming to customer specifications, electrical connections to Products manufactured by CSI, and Product specific training, is part of CSI’s product warranty. **CSI EXPRESSLY DISCLAIMS AND EXCLUDES ANY IMPLIED WARRANTIES OF MERCHANTABILITY OR FITNESS FOR A PARTICULAR PURPOSE.** CSI hereby disclaims, to the fullest extent allowed by applicable law, any and all warranties and conditions with respect to the Products, whether express, implied or statutory, other than those expressly provided herein.”

Assistance

Products may not be returned without prior authorization. The following contact information is for US and international customers residing in countries served by Campbell Scientific, Inc. directly. Affiliate companies handle repairs for customers within their territories. Please visit www.campbellsci.com to determine which Campbell Scientific company serves your country.

To obtain a Returned Materials Authorization (RMA), contact CAMPBELL SCIENTIFIC, INC., phone (435) 227-9000. After an application engineer determines the nature of the problem, an RMA number will be issued. Please write this number clearly on the outside of the shipping container. Campbell Scientific's shipping address is:

CAMPBELL SCIENTIFIC, INC.

RMA# _____
815 West 1800 North
Logan, Utah 84321-1784

For all returns, the customer must fill out a "Statement of Product Cleanliness and Decontamination" form and comply with the requirements specified in it. The form is available from our website at www.campbellsci.com/repair. A completed form must be either emailed to repair@campbellsci.com or faxed to (435) 227-9106. Campbell Scientific is unable to process any returns until we receive this form. If the form is not received within three days of product receipt or is incomplete, the product will be returned to the customer at the customer's expense. Campbell Scientific reserves the right to refuse service on products that were exposed to contaminants that may cause health or safety concerns for our employees.

Safety

DANGER — MANY HAZARDS ARE ASSOCIATED WITH INSTALLING, USING, MAINTAINING, AND WORKING ON OR AROUND TRIPODS, TOWERS, AND ANY ATTACHMENTS TO TRIPODS AND TOWERS SUCH AS SENSORS, CROSSARMS, ENCLOSURES, ANTENNAS, ETC. FAILURE TO PROPERLY AND COMPLETELY ASSEMBLE, INSTALL, OPERATE, USE, AND MAINTAIN TRIPODS, TOWERS, AND ATTACHMENTS, AND FAILURE TO HEED WARNINGS, INCREASES THE RISK OF DEATH, ACCIDENT, SERIOUS INJURY, PROPERTY DAMAGE, AND PRODUCT FAILURE. TAKE ALL REASONABLE PRECAUTIONS TO AVOID THESE HAZARDS. CHECK WITH YOUR ORGANIZATION'S SAFETY COORDINATOR (OR POLICY) FOR PROCEDURES AND REQUIRED PROTECTIVE EQUIPMENT PRIOR TO PERFORMING ANY WORK.

Use tripods, towers, and attachments to tripods and towers only for purposes for which they are designed. Do not exceed design limits. Be familiar and comply with all instructions provided in product manuals. Manuals are available at www.campbellsci.com or by telephoning (435) 227-9000 (USA). You are responsible for conformance with governing codes and regulations, including safety regulations, and the integrity and location of structures or land to which towers, tripods, and any attachments are attached. Installation sites should be evaluated and approved by a qualified engineer. If questions or concerns arise regarding installation, use, or maintenance of tripods, towers, attachments, or electrical connections, consult with a licensed and qualified engineer or electrician.

General

- Prior to performing site or installation work, obtain required approvals and permits. Comply with all governing structure-height regulations, such as those of the FAA in the USA.
- Use only qualified personnel for installation, use, and maintenance of tripods and towers, and any attachments to tripods and towers. The use of licensed and qualified contractors is highly recommended.
- Read all applicable instructions carefully and understand procedures thoroughly before beginning work.
- Wear a **hardhat** and **eye protection**, and take **other appropriate safety precautions** while working on or around tripods and towers.
- **Do not climb** tripods or towers at any time, and prohibit climbing by other persons. Take reasonable precautions to secure tripod and tower sites from trespassers.
- Use only manufacturer recommended parts, materials, and tools.

Utility and Electrical

- **You can be killed** or sustain serious bodily injury if the tripod, tower, or attachments you are installing, constructing, using, or maintaining, or a tool, stake, or anchor, come in **contact with overhead or underground utility lines**.
- Maintain a distance of at least one-and-one-half times structure height, 20 feet, or the distance required by applicable law, **whichever is greater**, between overhead utility lines and the structure (tripod, tower, attachments, or tools).
- Prior to performing site or installation work, inform all utility companies and have all underground utilities marked.
- Comply with all electrical codes. Electrical equipment and related grounding devices should be installed by a licensed and qualified electrician.

Elevated Work and Weather

- Exercise extreme caution when performing elevated work.
- Use appropriate equipment and safety practices.
- During installation and maintenance, keep tower and tripod sites clear of un-trained or non-essential personnel. Take precautions to prevent elevated tools and objects from dropping.
- Do not perform any work in inclement weather, including wind, rain, snow, lightning, etc.

Maintenance

- Periodically (at least yearly) check for wear and damage, including corrosion, stress cracks, frayed cables, loose cable clamps, cable tightness, etc. and take necessary corrective actions.
- Periodically (at least yearly) check electrical ground connections.

WHILE EVERY ATTEMPT IS MADE TO EMBODY THE HIGHEST DEGREE OF SAFETY IN ALL CAMPBELL SCIENTIFIC PRODUCTS, THE CUSTOMER ASSUMES ALL RISK FROM ANY INJURY RESULTING FROM IMPROPER INSTALLATION, USE, OR MAINTENANCE OF TRIPODS, TOWERS, OR ATTACHMENTS TO TRIPODS AND TOWERS SUCH AS SENSORS, CROSSARMS, ENCLOSURES, ANTENNAS, ETC.

Table of Contents

PDF viewers: These page numbers refer to the printed version of this document. Use the PDF reader bookmarks tab for links to specific sections.

1. Introduction	1
2. Precautions	1
3. Initial Inspection.....	1
4. QuickStart.....	2
5. Overview	4
6. Specifications.....	4
6.1 Temperature Measurement.....	5
6.2 Relative Humidity Measurement	5
7. Installation	6
7.1 Wiring to Datalogger	6
7.2 Datalogger Programming.....	6
7.3 SDI12Recorder() Instruction.....	7
7.4 Installation.....	7
7.4.1 Installation in a 41303-5A or 41303-5B 6-Plate Shield.....	8
7.4.2 Installation in a 41003-5 or 41003-5A 10-Plate Shield Using 6637 Collar Adapter	8
7.4.3 Installation in a RAD06 6-Plate Shield or RAD10 10-Plate Shield.....	8
7.4.4 Mount the Shield.....	9
8. Operation	10
8.1 Sensor Measurement	10
8.2 Measurements at Fast Scan Rates	11
8.3 Long Cables	11
8.4 Power Conservation	11
8.5 Measuring Multiple SDI-12 Sensors.....	11
9. Troubleshooting and Maintenance	12
9.1 Troubleshooting	12
9.2 Maintenance	12
9.3 Calibration.....	13
9.4 Sensor Element Replacement.....	13
10. Attributions and References	16

Appendices

A. Importing *Short Cut* Code Into *CRBasic Editor* ... A-1

B. Example Programs B-1

C. Environmental Performance..... C-1

C.1	Tests to Defined Standards	C-1
C.2	Exposure to Pollutants	C-1
C.3	Operating Range of RH Element.....	C-2
C.4	Measurement Below 0 °C.....	C-2

D. SDI-12 Sensor Support D-1

D.1	SDI-12 Command Basics	D-1
D.1.1	Address Query Command (!)	D-2
D.1.2	Change Address Command (aAb!)	D-2
D.1.3	Send Identification Command (aI!).....	D-2
D.1.4	Start Measurement Commands (aM!)	D-2
D.1.5	Aborting a Measurement Command	D-2
D.1.6	Send Data Command (aD!)	D-2
D.2	SDI-12 Transparent Mode.....	D-3
D.2.1	CR200(X) Series Datalogger Example	D-3
D.2.2	CR1000 Datalogger Example.....	D-4

Figures

7-1.	CS215 and 41303-5A Radiation Shield (left) and RAD06 Radiation Shield (right) on a tripod mast.....	9
7-2.	CS215 and 41303-5A Radiation Shield on a CM200 Series Crossarm	9
9-1.	Correct fit of sensor element (side view).....	15
9-2.	Incorrect fit of sensor element (side view)	16
C-1.	Normal operating conditions of RH element.....	C-2
D-1.	CR200(X) example of using the SDI-12 transparent mode to change the SDI-12 address from 0 to 1.....	D-4
D-2.	CR1000 example of using the SDI-12 transparent mode to change the SDI-12 address from 0 to 3. Sensor is connected to control port 7.....	D-5

Tables

7-1.	Wire Color, Function, and Datalogger Connection.....	6
C-1.	Environmental Tests.....	C-1
D-1.	SDI-12 Command and Response Set	D-1

CRBasic Examples

B-1.	Program Example for CR200(X) Series.....	B-1
B-2.	Program Example for CR300-Series, CR6-Series, CR800-Series, CR1000, CR3000, and CR5000 Dataloggers.....	B-2

CS215 Temperature and Relative Humidity Probe

1. Introduction

The CS215 Temperature and Relative Humidity probe is designed for general meteorological and other datalogging applications. It utilizes the SDI-12 communications protocol to communicate with any SDI-12 recorder, simplifying installation and programming.

Before using the CS215, please study:

- Section 2, *Precautions* (p. 1)
- Section 3, *Initial Inspection* (p. 1)
- Section 4, *QuickStart* (p. 2)

For Edlog datalogger support, check the availability of an older manual at www.campbellsci.com/old-manuals, or contact a Campbell Scientific application engineer for assistance.

2. Precautions

- READ AND UNDERSTAND the *Safety* section at the front of this manual.
- When opening the shipping package, do not damage or cut the cable jacket. If damage to the cable is suspected, consult with a Campbell Scientific application engineer.
- Although rugged, the CS215 should be handled as a precision scientific instrument.
- Santoprene® rubber, which composes the black outer jacket of the CS215 cable, will support combustion in air. It is used because of its resistance to temperature extremes, moisture, and UV degradation. It is rated as slow burning when tested according to U.L. 94 H.B. and passes FMVSS302. However, local fire codes may preclude its use inside buildings.

3. Initial Inspection


- Check the packaging and contents of the shipment. If damage occurred during transport, immediately file a claim with the carrier. Contact Campbell Scientific to facilitate repair or replacement.
- Check model information against the shipping documents to ensure the expected products and the correct lengths of cable are received. Model numbers are found on each product. On cables and cabled items, the model number is usually found at the connection end of the cable. Report any shortages immediately to Campbell Scientific.

4. QuickStart


Short Cut is an easy way to program your datalogger to measure the CS215 sensor and assign datalogger wiring terminals. *Short Cut* is available as a download on www.campbellsci.com and the *ResourceDVD*. It is included in installations of *LoggerNet*, *PC200W*, *PC400*, or *RTDAQ*.


Use the following procedure to get started.


1. Open *Short Cut*. Click **New Program**.




2. Select **Datalogger Model** and **Scan Interval** (default of 5 seconds is OK for most applications). Click **Next**.



- Under the **Available Sensors and Devices** list, select the **Sensors | Meteorological | Relative Humidity & Temperature** folder. Select **CS215 Temperature & Relative Humidity Sensor**. Click  to move the selection to the **Selected** device window. Temperature defaults to degree Celsius. This can be changed by clicking the **Deg C** box and selecting **Deg F**, for degrees Fahrenheit, or **K** for Kelvin. **SDI-12 Address** defaults to 0. Enter the correct **SDI-12 Address** for the CS215 if it has been changed from the factory-set default value.



- After selecting the sensor, click **Wiring Diagram** to see how the sensor is to be wired to the datalogger. The wiring diagram can be printed now or after more sensors are added.



5. Select any other sensors you have, then finish the remaining *Short Cut* steps to complete the program. The remaining steps are outlined in *Short Cut Help*, which is accessed by clicking on **Help | Contents | Programming Steps**.
6. If *LoggerNet*, *PC400*, *RTDAQ*, or *PC200W* is running on your PC, and the PC to datalogger connection is active, you can click **Finish** in *Short Cut* and you will be prompted to send the program just created to the datalogger.
7. If the sensor is connected to the datalogger, as shown in the wiring diagram in step 4, check the output of the sensor in the datalogger support software data display to make sure it is making reasonable measurements.

5. Overview

The CS215 probe uses a single chip element that incorporates both a temperature and an RH sensor. Each element is individually calibrated with the calibration corrections stored on the chip. The element is easily changed in the field, reducing downtime and calibration costs.

Electronics within the CS215 control the measurement made by the sensor element, apply temperature and linearization corrections to the readings, and present the data via SDI-12 to a datalogger.

An inner expanded PTFE filter minimizes the effects of dust and dirt on the sensor. The filter is lightweight and hydrophobic, thereby diminishing its effect on the time response of the sensor.

The probe housing is designed to withstand permanent exposure to all weather and to fit into a range of radiation shields, including compact shields.

6. Specifications

Features:

- Accurate, stable measurements
- Field-changeable element allows on-site recalibration
- Individually calibrated sensor elements require no further adjustment of the probe
- Low power consumption
- Digital SDI-12 output
- Compatible with Campbell Scientific dataloggers CR200(X) series, CR300 series, CR6 series, CR800 series, CR1000, CR3000, and CR5000.

Sensor Element: Sensirion SHT75

Calibration Traceability: NIST and NPL standards

Supply Voltage: 7 to 28 Vdc for serial numbers E13405 and newer
6 to 18 Vdc for older models

Current Consumption: 70 μ A quiescent, typical
1.7 mA during 0.7 s measurement

Diameter:	12 mm at sensor tip, maximum 18 mm at cable end
Length:	180 mm, including cable strain relief
Housing Material:	Anodized aluminum
Housing Classification:	IP65 (NEMA 4)
Filter Material:	Inner expanded PTFE filter. Filter material has a porosity of 64% and a pore size of < 3µm.
EMC Compliance:	Tested and conforms to IEC61326:2002

6.1 Temperature Measurement

Operating Range:	−40 to +70 °C
Accuracy:	±0.3 °C at 25 °C ±0.4 °C over 5 to 40 °C ±0.9 °C over −40 to 70 °C
Response Time with Filter:	120 s (63% response time in air moving at 1 m/s)
Default Units:	Degrees Celsius

6.2 Relative Humidity Measurement

Operating Range:	0 to 100% RH (−20 to 60 °C; see Appendix C, <i>Environmental Performance (p. C-1)</i>)
Accuracy at 25 °C:	±2% over 10 to 90% ±4% over 0 to 100%
Short-Term Hysteresis:	<1% RH
Temperature Dependence:	Compensated to better than ±2% over −20 to 60 °C
Typical Long-Term Stability:	Better than ±1.0% per year
Response Time with Filter:	<10 s (63% response time in air moving at 1 m/s at humidity <85%)
Environmental Performance:	See Appendix C, <i>Environmental Performance (p. C-1)</i>

7. Installation

If you are programming your datalogger with *Short Cut*, skip Section 7.1, *Wiring to Datalogger* (p. 6), and Section 7.2, *Datalogger Programming* (p. 6). *Short Cut* does this work for you. See Section 4, *QuickStart* (p. 2), for a *Short Cut* tutorial.

7.1 Wiring to Datalogger

TABLE 7-1. Wire Color, Function, and Datalogger Connection		
Wire Color	Wire Function	Datalogger Connection Terminal
Red	Power	12V
White	Power ground	G
Black	Power ground	G
Green	SDI-12 signal	Control Port ¹ or U configured for SDI-12 ²
Clear	Shield	⏏
¹ Dedicated SDI-12 port on CR5000		
² U channels are automatically configured by the measurement instruction.		

To use more than one probe per datalogger, either connect the different probes to different terminals on the datalogger or change the SDI-12 addresses of the probes and wire them to the same terminal. Using the SDI-12 address reduces the use of ports on the datalogger and allows probes to be connected in a “daisy-chain” fashion which can minimize cable runs in some applications. (See below for limits on the total cable length.)

The SDI-12 address of the CS215 can be set two ways:

- By sending the required commands to the sensors via an SDI-12 recorder/datalogger that allows talk through to the sensor.
- By loading a program into the recorder that sends the required commands

See Appendix D, *SDI-12 Sensor Support* (p. D-1), for detailed instructions.

7.2 Datalogger Programming

Short Cut is the best source for up-to-date datalogger programming code. Programming code is needed when:

- Creating a program for a new datalogger installation
- Adding sensors to an existing datalogger program

If your data acquisition requirements are simple, you can probably create and maintain a datalogger program exclusively with *Short Cut*. If your data acquisition needs are more complex, the files that *Short Cut* creates are a great source for programming code to start a new program or add to an existing custom program.

NOTE

Short Cut cannot edit programs after they are imported and edited in *CRBasic Editor*.

A *Short Cut* tutorial is available in Section 4, *QuickStart* (p. 2). If you wish to import *Short Cut* code into *CRBasic Editor* to create or add to a customized program, follow the procedure in Appendix A, *Importing Short Cut Code Into CRBasic Editor* (p. A-1). Programming basics for CRBasic dataloggers are provided in the following sections. Complete program examples for select dataloggers can be found in Appendix B, *Example Programs* (p. B-1).

7.3 SDI12Recorder() Instruction

The **SDI12Recorder()** measurement instruction programs CRBasic dataloggers (CR200(X)-, CR300-, CR6-, and CR800-series, CR1000, CR3000, and CR5000) to measure the CS215 sensor. This instruction sends a request to the sensor to make a measurement and then retrieves the measurement from the sensor. See Section 8.1, *Sensor Measurement* (p. 10), for more information.

SDI12Recorder(Dest, SDIPort, SDIAddress, "SDICommand", Mult, Offset)

The Destination parameter must be an array of length 2, with the first index for Air Temperature (in °C) and the second for Relative Humidity (as a percent).

Variations:

- Set SDICommand to "M!", "C!", or "R!" – see Section 8.1, *Sensor Measurement* (p. 10), to determine which is best for your application.

7.4 Installation

Locate the sensor over an open, level area at least 9 m (EPA) in diameter. The surface should be covered by short grass or the natural earth surface where grass does not grow. Sensors should be located at a distance of at least four times the height of any nearby obstruction and at least 30 m (EPA) from large, paved areas. Sensors should be protected from thermal radiation and adequately ventilated. Protect the filter at the top of the sensor from exposure to liquid water. The hydrophobic nature of the filter repels light rain, but driving rain can force itself into the pore structure of the filter and take time to dry out.

Standard measurement heights:

- 1.5 m (AASC)
- 1.25 to 2.0 m (WMO)
- 2.0 m (EPA)

See Section 10, *Attributions and References* (p. 16), for a list of references that discuss temperature and relative humidity sensors.

When used in the field, the CS215 must be housed in a radiation shield. Typically, the 41303-5A or RAD06 six-plate solar radiation shield is used.

A 41003-5 or RAD10 ten-plate solar radiation shield can also house the CS215 sensor. Using a 10-plate shield will give slightly more accurate readings. The

41003-5 shield will require a special adapter (pn 6637). The RAD10 will not require any additional parts.

The white color of these shields reflects solar radiation, and the louvered construction allows air to pass freely through, thereby keeping the probe at or near ambient temperature. The RAD06 and RAD10 use a double-louvered design that offers improved sensor protection from insect intrusion and driving rain and snow. In addition, the RAD06 and RAD10 shields have lower self-heating in bright sunlight combined with higher temperatures ($> 24^{\circ}\text{C}$ (75°F)) and low wind speeds ($< 2\text{ m/s}$ (4.5 mph)), giving a better measurement.

Each of these solar radiation shields attaches to a crossarm, mast, or user-supplied pipe with a 2.5 to 5.3 cm (1.0 to 2.1 inch) outer diameter.

Tools required for installing a radiation shield to a tripod or tower include:

- 1/2-inch open-end wrench
- small screwdriver provided with datalogger
- small Phillips screwdriver
- UV-resistant cable ties
- small pair of diagonal-cutting pliers
- adjustable wrench with a minimum 1-7/8 inch jaw size

7.4.1 Installation in a 41303-5A or 41303-5B 6-Plate Shield

1. With a small Phillips screwdriver, loosen the plastic split collar at the base of the shield (reversing the removable portion if necessary), and gently insert the probe.
2. Tighten the screws on the collar until it firmly grips the probe body. See [FIGURE 7-1](#) (left) and [FIGURE 7-2](#).

7.4.2 Installation in a 41003-5 or 41003-5A 10-Plate Shield Using 6637 Collar Adapter

1. Loosely thread the collar adapter into the base of the 10-plate shield.
2. Insert the CS215 sensor through the collar as far as it will go.
3. Hold the collar and sensor, and finish threading the collar into the shield by hand. Tighten the collar around the sensor until it firmly grips the probe body. Use an adjustable wrench if necessary, but do not overtighten the collar.

7.4.3 Installation in a RAD06 6-Plate Shield or RAD10 10-Plate Shield

1. Loosen the nut on the entry gland at the bottom of the shield.
2. Insert the sensor into the gland as far as it will go. See [FIGURE 7-1](#) (right).
3. Using an adjustable wrench, tighten the nut on the gland until the sensor is held firmly in place. Do not overtighten.

7.4.4 Mount the Shield

1. Attach the radiation shield to the tripod mast, crossarm, or tower leg using the supplied U-bolt or band clamp. See FIGURE 7-1 and FIGURE 7-2 for examples of shield mounting.
2. Route the cable to the datalogger, and secure the cable to the mounting structure using cable ties.

CAUTION

Failure to secure the cable can lead to breakage of the wires due to fatigue caused by blowing back and forth in the wind.



FIGURE 7-1. CS215 and 41303-5A Radiation Shield (left) and RAD06 Radiation Shield (right) on a tripod mast



FIGURE 7-2. CS215 and 41303-5A Radiation Shield on a CM200 Series Crossarm

8. Operation

8.1 Sensor Measurement

CRBasic instruction **SDI12Recorder()** measures the CS215 sensor using the Start Measurement (**aM!**), Concurrent Measurement (**aC!**), or Continuous Measurement (**aR!**) command. In all cases “a” is the address of the sensor and “!” is the command terminator. These two characters are sent implicitly by all CRBasic dataloggers but the CR200(X) series. The sensor returns two values: temperature in degrees Celsius and relative humidity as a percentage (0 to 100).

The factory-set SDI-12 address for the CS215 is 0. See Appendix [D.1.2](#), *Change Address Command* (**aAb!**) (*p. D-2*), for instructions on changing the SDI-12 address through the datalogger.

- A measurement is initiated with the **aM!** or **aC!** command. To these commands, the sensor responds with the time until the measurement data are available and the number of values to be returned when one or more subsequent **aD!** commands are issued.

When using the **aM!** command, the datalogger waits for the time specified by the sensor, sends the **aD!** command, pauses its operation, and waits until either it receives the data from the sensor or the sensor timeout expires. If the datalogger receives no response, it will send the command a total of three times, with three retries for each attempt, or until a response is received. Because of the delays this command requires, it is only recommended in measurement scans of 10 seconds or more.

- The **aC!** command follows the same pattern as the **aM!** command with the exception that it does not require the datalogger to pause its operation until the values are ready. Rather, the datalogger picks up the data with the **aD!** command on the next pass through the program. Another measurement request is then sent so that data are ready on the next scan.
- An **aR!** command switches the sensor to automatically make measurements and send data every 11 seconds, ± 2 seconds, based on the sensor's internal clock. If measurements are requested at 2 seconds or faster, the sensor will increase its measurement rate to approximately every 5 seconds. This instruction usually takes less than 300 milliseconds to execute. The automatic measurement mode can only be cancelled by powering down the sensor to reset it.

NOTE

Only CS215 sensors with serial numbers higher than E1587 or those with upgraded firmware support the **aR!** command.

The CS215 also supports the **aMC!**, **aCC!**, and **aRC!** instructions, which are the same as the instructions above, but where the C at the end of the instruction forces a validation for the data received from the sensor using a checksum. If the checksum is invalid, the datalogger will re-request the data up to three times. The checksum validation increases the measurement time by about 40 milliseconds if there are no errors. Retries will increase the measurement time

in proportion to the number of retries. Use of the checksum option is only normally necessary for very long cable runs.

See Appendix D, *SDI-12 Sensor Support (p. D-1)*, for additional commands and details of the SDI-12 protocol.

8.2 Measurements at Fast Scan Rates

Using the **SlowSequence()** function allows the SDI-12 instruction to run as a background process, causing minimum interference to other measurements that use the analog hardware. Measuring the sensor in a **SlowSequence()** section of the program allows faster programs to run as the main scan.

NOTE

For the CR5000, use a control port rather than the SDI-12 port to allow the SDI12recorder instruction to run in the slow sequence.

8.3 Long Cables

Digital data transfer eliminates offset errors due to cable lengths. However, digital communications can break down when cables are too long, resulting in either no response from the sensor or corrupted readings. The original SDI-12 standard specifies the maximum total cable length to be 200 feet (61 meters). To ensure proper operation with long cables, follow these guidelines:

- Use low capacitance, low resistance, screened cable (as fitted by Campbell Scientific) to reach distances of several hundred meters.
- Ensure that the power ground cable has low resistance and is connected to the same ground reference as the datalogger control ports.
- Be aware that “daisy-chaining” sensors reduces the maximum cable length roughly in proportion to the number of sensors connected in parallel.

8.4 Power Conservation

The CS215 draws less than 70 μA of current between measurements. In most applications this is insignificant compared to the datalogger and other power draws, so the sensor can be permanently connected.

In low-power applications, conserve battery power by turning the 12 V supply to the CS215 on just before the measurement (allowing a warm-up time of at least 100 ms) and then turning it off afterwards. If available, the switched 12 V output of the datalogger can be used.

8.5 Measuring Multiple SDI-12 Sensors

Up to ten CS215s or a combination of other SDI-12 sensors can be connected to a single datalogger control port. Each SDI-12 device must have a unique SDI-12 address between 0-9, A-Z, or a-z. See Appendix D.1.2, *Change Address Command (aAb!)* (p. D-2) to change the CS215 SDI-12 address from its default address of 0.

9. Troubleshooting and Maintenance

NOTE

All factory repairs and recalibrations require a returned material authorization (RMA) and completion of the “Declaration of Hazardous Material and Decontamination” form. Refer to the [Assistance](#) page at the beginning of this manual for more information.

9.1 Troubleshooting

Symptom: Temperature is reported as **–9999** or **NAN**, and relative humidity is reported as **0** or as an unchanging value.

Recheck wiring. Verify the green wire is connected to the control port specified by the **SDI12Recorder()** instruction. Verify the red wire is connected to a 12V terminal.

Check the voltage to the sensor with a digital voltage meter. If a switched 12V terminal is used, temporarily connect the red wire to a 12V terminal (non-switched) for test purposes.

Verify the probe’s SDI12 address matches the address entered for the **SDI12Recorder()** instruction. The default address is 0. The address can be verified or changed with the commands described in Appendix D, *SDI-12 Sensor Support* (p. D-1).

Remove the filter tip and verify that the sensing element has been installed with the proper orientation as described in Section 9.4, *Sensor Element Replacement* (p. 13).

Symptom: Incorrect temperature or relative humidity is reported.

If using the SW12 terminal to power the sensor, verify the program is allowing a warm-up time of at least 100 ms.

Check to see if the filter tip has been contaminated. Replace the filter tip (pn 18142), or clean with distilled water as needed.

9.2 Maintenance

The CS215 probe requires minimal maintenance, but dust, debris, and salts on the filter cap will degrade sensor performance. Check the white filter on the end of the sensor for debris. If dirt or salt is ingrained in the filter, clean with distilled water or replace it. Make sure the filter is connected firmly with your fingers—do not over tighten

Check the radiation shield monthly to make sure it is free from dust and debris. To clean the shield, first remove the sensor. Dismount the shield. Brush all loose dirt off. If more effort is needed, use warm, soapy water and a soft cloth or brush to thoroughly clean the shield. Allow the shield to dry before remounting.

9.3 Calibration

The life of the humidity chip element is quoted as many years with a typical drift of less than 1% per year when used in ‘clean’ environments. Because it can be difficult to know what the sensor has been exposed to and because the element is relatively inexpensive, we recommend replacing the sensor element at the normal interval you would recalibrate similar probes, e.g., annually. Replacing the element effectively brings the probe back to a factory calibration state both for temperature and relative humidity.

To check an old element’s calibration in order to formally record the probe’s pre-calibration state (pre-sensor replacement), you can measure its performance by plugging it into another sensor.

9.4 Sensor Element Replacement

To replace the element:

1. Wash your hands to avoid getting dirt or grease on the element.

NOTE

It is virtually impossible to touch and damage the sensing elements which are enclosed within the black molded plastic at the tip. However, if dirt, salt or grease are left on the plastic during the process of handling the element, it may influence the measurements made.

2. Disconnect the sensor from the 12 V power supply.
3. Remove the filter by unscrewing it counterclockwise when looking towards the tip of the sensor.

CAUTION

The filter cap unscrews from the probe. Attempting to pull it off will destroy it.

4. Identify the sensor element. FIGURE 9-1 below shows a side view of the end of the probe and sensor element. Before removing the element, carefully study the probe, note its orientation, and read the following description:
 - The element plugs into the black plastic socket that protrudes by about 1 mm from the end of the metal body of the sensor.
 - There are eight holes in the socket, while the element only has four pins.
 - The element will work when fitted into either side of socket but must be installed in one of the two possible orientations to work.
 - The correct orientation is with the black molded tip of the element (that contains the sensing components) mounted directly above the center of the socket.
 - FIGURE 9-1 shows the correct orientation, while FIGURE 9-2 shows the incorrect orientation.

CAUTION

If the sensor is oriented incorrectly, it will not work. The element will draw excessive power from the supply and may be damaged if left powered in this state for more than a few seconds.

5. Grasp the body of the sensor (this also ensures you are at the same electrical potential as the element) and, holding the black tip of the element between your fingertips, pull the element out of the socket. Store the old element in electrostatic protective packaging if you wish to retain it.
6. With the element removed, check for dirt and/or corrosion around the socket. Clean any dirt away using a damp cloth to remove any salts that might be there.
7. Unpack the replacement element, avoiding static discharges to the element by making sure you touch the packaging before the element.
8. Either hold the element by the black top of the package (the other end to the gold plated pins) or use a pair of fine nosed pliers or tweezers to grip the sensor by the pins. Carefully match the pins to the socket in the end of probe.
9. Confirm the correct orientation and gently push the pins into the socket until they will not go in any further.
10. Before replacing the filter element and turning on power to the sensor, double-check that the sensor is inserted in the correct orientation. Refer to [FIGURE 9-1](#).
11. Screw the filter back onto the end of the probe, making sure it clears the sensor element. If the element appears too close to the filter, there is a fair chance that it has been inserted in the incorrect orientation or that the legs of the element have been bent. Screw the filter onto the thread and tighten gently with your fingers.

CAUTION

Only tighten the filter approximately 1/8th of a turn by hand when the filter is fully screwed onto the thread. Over-tightening the filter will damage it and cause problems in inserting and removing the probe from some shields.




FIGURE 9-1. Correct fit of sensor element (side view)




FIGURE 9-2. Incorrect fit of sensor element (side view)

10. Attributions and References

Santoprene® is a registered trademark of Exxon Mobile Corporation.

AASC, 1985: The State Climatologist (1985) Publication of the American Association of State Climatologists: *Heights and Exposure Standards for Sensors on Automated Weather Stations*, v. 9, No. 4 October, 1985. (www.stateclimate.org/publications/state-climatologist/NOAA-NCY-SCBOOKS-SC77097/00000029.pdf)

EPA, 2008: *Quality Assurance Handbook for Air Pollution Measurement Systems*, Vol. IV, Meteorological Measurements, Ver. 2.0, EPA-454/B-08-002 (revised 2008). Office of Air Quality Planning and Standards, Research Triangle Park, NC 27711.

Goff, J. A. and S. Gratch, 1946: Low-pressure properties of water from -160° to 212°F, *Trans. Amer. Soc. Heat. Vent. Eng.*, **51**, 125-164.

Lowe, P. R., 1977: An approximating polynomial for the computation of saturation vapor pressure, *J. Appl. Meteor.*, **16**, 100-103.

Meyer, S. J. and K. G. Hubbard, 1992: Nonfederal Automated Weather Stations and Networks in the United States and Canada: A Preliminary Survey, *Bulletin Am. Meteor. Soc.*, **73**, No. 4, 449-457.

Weiss, A., 1977: Algorithms for the calculation of moist air properties on a hand calculator, *Amer. Soc. Ag. Eng.*, **20**, 1133-1136.

WMO, 2008. *Guide to Meteorological Instruments and Methods of Observation*. World Meteorological Organization No. 8, 7th edition, Geneva, Switzerland.

Appendix A. Importing Short Cut Code Into CRBasic Editor

This tutorial shows:

- How to import a *Short Cut* program into a program editor for additional refinement
- How to import a wiring diagram from *Short Cut* into the comments of a custom program

Short Cut creates files, which can be imported into *CRBasic Editor*. Assuming defaults were used when *Short Cut* was installed, these files reside in the C:\campbellsci\SCWin folder:

- .DEF (wiring and memory usage information)
- .CR2 (CR200(X)-series datalogger code)
- .CR300 (CR300-series datalogger code)
- .CR6 (CR6-series datalogger code)
- .CR8 (CR800-series datalogger code)
- .CR1 (CR1000 datalogger code)
- .CR3 (CR3000 datalogger code)
- .CR5 (CR5000 datalogger code)

Use the following procedure to import *Short Cut* code and wiring diagram into *CRBasic Editor*.

1. Create the *Short Cut* program following the procedure in Section 4, *QuickStart* (p. 2). Finish the program and exit *Short Cut*. Make note of the file name used when saving the *Short Cut* program.
2. Open *CRBasic Editor*.
3. Click **File | Open**. Assuming the default paths were used when *Short Cut* was installed, navigate to C:\CampbellSci\SCWin folder. The file of interest has the .CR2, .CR300, .CR6, .CR8, .CR1, .CR3, or .CR5 extension. Select the file and click **Open**.
4. Immediately save the file in a folder different from C:\Campbellsci\SCWin, or save the file with a different file name.

NOTE

Once the file is edited with *CRBasic Editor*, *Short Cut* can no longer be used to edit the datalogger program. Change the name of the program file or move it, or *Short Cut* may overwrite it next time it is used.

5. The program can now be edited, saved, and sent to the datalogger.
6. Import wiring information to the program by opening the associated .DEF file. Copy and paste the section beginning with heading “-Wiring for CRXXX-” into the CRBasic program, usually at the head of the file. After pasting, edit the information such that an apostrophe (') begins each line. This character instructs the datalogger compiler to ignore the line when compiling.

Appendix B. Example Programs

This example program shows the measurement of a single CS215 and can be used directly with CR200(X) series dataloggers.

CRBasic Example B-1. Program Example for CR200(X) Series

```
'CR200(X) Series Datalogger
'Program measures one CS215 sensor every 30 seconds and stores the average
'temperature and a sample of relative humidity every 10 minutes.

'Wiring Diagram
'=====
'CS215

' Wire
' Color   Function      CR200(X)
' -----
' Red     Power (12V)    Battery +
' Green   SDI-12 signal  C1/SDI-12
' Black   Power ground   G
' White   Power ground   G
' Clear   Shield         Signal Ground

'Declare the variable array for the measurement
Public TRHData(2)

Alias TRHData(1)=AirTC
Alias TRHData(2)=RH

Units AirTC=Deg C
Units RH=%

'Define a data table for ten-minute data
DataTable(TenMin,True,-1)
  DataInterval(0,10,Min)
  Average(1,AirTC,False)
  Sample(1,RH)
EndTable

'Main Program
BeginProg
  Scan (30,Sec) 'Scan every 30 seconds
  'CS215 Temperature & Relative Humidity Sensor measurements 'AirTC' and 'RH'
  SDI12Recorder(TRHData(),"OM!",1,0)
  'Call Data Tables and Store Data
  CallTable TenMin
  NextScan
EndProg
```

This program can be adapted for use with CR300-, CR6-, and CR800-series, CR1000, CR3000, and CR5000 dataloggers.

CRBasic Example B-2. Program Example for CR300-Series, CR6-Series, CR800-Series, CR1000, CR3000, and CR5000 Dataloggers

```
'Program measures one CS215 sensor every 5 seconds and stores the average
'temperature and a sample of relative humidity every 10 minutes.

'Wiring Diagram
'=====
'CS215

' Wire
' Color      Function      CR1000
' -----
' Red        Power (12V)    12V
' Green      SDI-12 signal  C7
' Black      Power ground   G
' White      Power ground   G
' Clear      Shield         Signal Ground

'Declare the variable array for the measurement
Public TRHData(2)

Alias TRHData(1)=AirTC
Alias TRHData(2)=RH

Units AirTC=Deg C
Units RH=%

'Define Data Tables
DataTable(TenMin,True,-1)
  DataInterval(0,10,Min,10)
  Average(1,AirTC,FP2,False)
  Sample(1,RH,FP2)
EndTable

'Main Program
BeginProg
'Main Scan
Scan(5,Sec,1,0)
  'CS215 Temperature & Relative Humidity Sensor measurements 'AirTC' and 'RH'
  SDI12Recorder(TRHData(),7,"0","M!",1,0)
  'Call Data Tables and Store Data
  CallTable(TenMin)
NextScan
EndProg
```

Appendix C. Environmental Performance

This Appendix details tests and limitations of the sensor when exposed to extremes of the environment.

C.1 Tests to Defined Standards

The sensor element has been tested by the manufacturer and found to comply with various environmental test standards as shown in TABLE C-1 below:

TABLE C-1. Environmental Tests		
Environment	Norm	Results
Temperature Cycles	JESD22-A104-B –40/+125 °C, 1000 cycles	Within Specifications
HAST Pressure Cooker	JESD22-A110-B 2.3bar 125 °C 85% RH	Reversible shift by +2% RH
Salt Atmosphere	DIN-50021SS	Within Specifications
Condensing Air	–	Within Specifications
Freezing cycles Fully Submerged	–20/+90 °C, 100 cycles, 30 min dwell time	Reversible shift by +2% RH
Various Automotive Chemicals	DIN 72300-5	Within Specifications
Cigarette Smoke	Equivalent to 15years in a mid-size car	Within Specifications

N.B. The temperature sensor passed all tests without any detectable drift. Package and electronics also passed 100%

C.2 Exposure to Pollutants

All capacitive sensors are susceptible to pollutants to some degree. The vapors may interfere with the polymer layers used in the structure of the sensing element. The diffusion of chemicals into the polymer may cause temporary or even permanent shifts in both offset and sensitivity.

After low levels of exposure, in a clean environment the contaminants will slowly outgas and the sensor recovers. High levels of pollutants may cause permanent damage to the sensing polymer.

As a general rule, the sensor will not be damaged by levels of chemicals which are not too dangerous to human health (see TABLE C-1), so damage is not normally a problem in outdoor applications. Avoid exposing the sensor to chemicals at higher concentrations.

C.3 Operating Range of RH Element

The RH sensor is specified to work over the entire humidity range of 0–100% RH for the temperature range –20 to 60 °C. It will give readings over an extended range as shown in FIGURE C-1 below (although the electronics of the CS215 probe are not specified to operate beyond 70 °C).

When used outside the range of normal conditions or when subject to prolonged periods of condensation or freezing, the sensor calibration may be temporarily altered, normally resulting in a change of <+3% RH. Upon returning to normal conditions, the calibration will settle back to the “standard” calibration over the course of several days. In laboratory conditions, it is possible to speed up this process by a reconditioning process, as follows: 80–90 °C at < 5 %RH for 24 h (baking) followed by 20–30 °C at > 74% RH for 48 h (re-hydration).




FIGURE C-1. Normal operating conditions of RH element

C.4 Measurement Below 0 °C

The CS215 provides a humidity reading that is referenced to the saturated water vapor pressure above liquid water, even at temperatures below 0 °C, where ice might form. This is the common way to express relative humidity and is as defined by the World Meteorological Organization. If an RH value is required to be referenced to ice, the CS215 readings will need to be corrected.

One consequence of using water as the reference is that the maximum humidity that will normally be output by the sensor for temperatures below freezing is as follows:

100% RH at 0 °C
 95% RH at –5 °C
 91% RH at –10 °C
 87% RH at –15 °C
 82% RH at –20 °C
 78% RH at –25 °C
 75% RH at –30 °C

In practical terms this means that, for instance, at –20 °C the air is effectively fully saturated when the sensor outputs 82% RH.

Appendix D. SDI-12 Sensor Support

D.1 SDI-12 Command Basics

SDI-12 commands have three components:

Sensor address (a) – a single character, and is the first character of the command. The default address of zero (0) can be used unless multiple sensors are connected to the same port.

Command body (e.g., M1) – an upper case letter (the “command”) followed by alphanumeric qualifiers.

Command termination (!) – an exclamation mark.

An active sensor responds to each command. Responses have several standard forms and terminate with <CR><LF> (carriage return – line feed). Standard SDI-12 commands supported by the CS215 are listed in TABLE D-1.

TABLE D-1. SDI-12 Command and Response Set		
Name	Command	Response
Acknowledge Active	a!	a<CR><LF>
Send Identification	aI!	allccccccmmmmmmmvvxxx...xx<CR><LF>
Change Address	aAb!	b<CR><LF>
Address Query	?!	a<CR><LF>
Start Measurement	aM!	atttn<CR><LF>
Start Measurement and Request CRC	aMC!	atttn <CR><LF>
Send Data	aD0!	a<values><CR><LF>
Start Concurrent Measurement	aC!	atttn<CR><LF>
Start Concurrent Measurement and Request CRC	aCC!	atttn <CR><LF>
Continuous Measurements	aR0!	a<values><CR><LF>

Continuous Measurements and Request CRC	aRC0!	a<values><CR><LF>
Start Verification	aV!	atttn

D.1.1 Address Query Command (?!)

Command **?!** requests the address of the connected sensor. The sensor replies to the query with the address, *a*.

D.1.2 Change Address Command (aAb!)

Sensor address is changed with command **aAb!**, where *a* is the current address and *b* is the new address. For example, to change an address from 0 to 2, the command is **0A2!**. The sensor responds with the new address *b*, which in this case is 2.

D.1.3 Send Identification Command (aI!)

Sensor identifiers are requested by issuing command **aI!**. The reply is defined by the sensor manufacturer, but usually includes the sensor address, SDI-12 version, manufacturer's name, and sensor model information. Serial number or other sensor specific information may also be included.

D.1.4 Start Measurement Commands (aM!)

A measurement is initiated with **M!** commands. The response to each command has the form atttn, where

a = sensor address

ttt = time, in seconds, until measurement data are available

nn = the number of values to be returned when one or more subsequent **D!** commands are issued.

D.1.5 Aborting a Measurement Command

A measurement command (**M!**) is aborted when any other valid command is sent to the sensor.

D.1.6 Send Data Command (aD!)

This command requests data from the sensor. It is normally issued automatically by the datalogger after measurement commands **aM!**. In transparent mode, the user asserts this command to obtain data.

D.2 SDI-12 Transparent Mode

System operators can manually interrogate and enter settings in probes using transparent mode. Transparent mode is useful in troubleshooting SDI-12 systems because it allows direct communication with probes. Datalogger security may need to be unlocked before transparent mode can be activated.

Transparent mode is entered while the PC is in telecommunications with the datalogger through a terminal emulator program. It is easily accessed through Campbell Scientific datalogger support software. Datalogger keyboards and displays cannot be used.

The terminal emulator is accessed by navigating to the Datalogger menu in *PC200W*, the Tools menu in *PC400*, or the Datalogger menu in the *Connect* screen of *LoggerNet*.

The following examples show how to use *LoggerNet* software to enter transparent mode and change the SDI-12 address of a CS215. The same steps are used to enter transparent mode with *PC200W* and *PC400* software after accessing the terminal emulator as previously described.

D.2.1 CR200(X) Series Datalogger Example

1. Connect a single CS215 to the CR200(X) (see TABLE 7-1).
2. In the *LoggerNet Connect* screen navigate to the **Datalogger** menu and select **Terminal Emulator**. The terminal emulator window will open. In the **Select Device** menu, located in the lower left-hand side of the window, select the CR200Series station.
3. Click on the **Open Terminal** button.
4. Press the <enter> key until the datalogger responds with the **CR2XX>** prompt. At the **CR2XX>** prompt, make sure the **All Caps Mode** box is checked, and enter the command **SDI12 <enter>**. The response **SDI12>** indicates that the CS215 is ready to accept SDI-12 commands.
5. To query the CS215 for its current SDI-12 address, key in **?! <enter>** and the CS215 will respond with its SDI-12 address. If no characters are typed within 60 seconds, then the mode is exited. In that case, simply enter the command **SDI12** again and press <enter>.
6. To change the SDI-12 address, key in **aAb! <enter>**, where *a* is the current address from the above step and *b* is the new address (see FIGURE D-1). The CS215 will change its address and the datalogger will respond with the new address. To exit SDI-12 transparent mode select the **Close Terminal** button.




FIGURE D-1. CR200(X) example of using the SDI-12 transparent mode to change the SDI-12 address from 0 to 1.

D.2.2 CR1000 Datalogger Example

1. Connect a CS215 to the CR1000 (see TABLE 7-1).
2. In the *LoggerNet Connect* screen, navigate to the **Datalogger** menu and select **Terminal Emulator**. The terminal emulator window will open. In the **Select Device** menu, located in the lower left-hand side of the window, select the CR1000 station.
3. Click on the **Open Terminal** button.
4. Press the <enter> key until the datalogger responds with the **CR1000>** prompt. At the **CR1000>** prompt, make sure the **All Caps Mode** box is checked and enter the command **SDI12** <enter>. In response, the datalogger will list all possible SDI-12 ports. At the **Select SDI12 Port:** prompt, key in the number of the control port where the CS215 is connected and press <enter>. The response **Entering SDI12 Terminal** indicates that the CS215 is ready to accept SDI-12 commands.
5. To query the CS215 for its current SDI-12 address, key in **?! <enter>** and the CS215 will respond with its SDI-12 address. If no characters are typed within 60 seconds, the mode is exited. In that case, simply enter the command **SDI12** again, press <enter>, and key in the correct control port number option when prompted.
6. To change the SDI-12 address, key in **aAb! <enter>**, where *a* is the current address from the above step and *b* is the new address (see FIGURE D-2). The CS215 will change its address and the datalogger will respond with the new address. To exit SDI-12 transparent mode, select the **Close Terminal** button.




FIGURE D-2. CR1000 example of using the SDI-12 transparent mode to change the SDI-12 address from 0 to 3. Sensor is connected to control port 7.

Campbell Scientific Companies

Campbell Scientific, Inc.

815 West 1800 North
Logan, Utah 84321
UNITED STATES

www.campbellsci.com • info@campbellsci.com

Campbell Scientific Canada Corp.

14532 – 131 Avenue NW
Edmonton AB T5L 4X4
CANADA

www.campbellsci.ca • dataloggers@campbellsci.ca

Campbell Scientific Africa Pty. Ltd.

PO Box 2450
Somerset West 7129
SOUTH AFRICA

www.campbellsci.co.za • cleroux@csafrica.co.za

Campbell Scientific Centro Caribe S.A.

300 N Cementerio, Edificio Breller
Santo Domingo, Heredia 40305
COSTA RICA

www.campbellsci.cc • info@campbellsci.cc

Campbell Scientific Southeast Asia Co., Ltd.

877/22 Nirvana@Work, Rama 9 Road
Suan Luang Subdistrict, Suan Luang District
Bangkok 10250
THAILAND

www.campbellsci.asia • info@campbellsci.asia

Campbell Scientific Ltd.

Campbell Park
80 Hathern Road
Shepshed, Loughborough LE12 9GX
UNITED KINGDOM

www.campbellsci.co.uk • sales@campbellsci.co.uk

Campbell Scientific Australia Pty. Ltd.

PO Box 8108
Garbutt Post Shop QLD 4814
AUSTRALIA

www.campbellsci.com.au • info@campbellsci.com.au

Campbell Scientific Ltd.

3 Avenue de la Division Leclerc
92160 ANTONY
FRANCE

www.campbellsci.fr • info@campbellsci.fr

Campbell Scientific (Beijing) Co., Ltd.

8B16, Floor 8 Tower B, Hanwei Plaza
7 Guanghua Road
Chaoyang, Beijing 100004
P.R. CHINA

www.campbellsci.com • info@campbellsci.com.cn

Campbell Scientific Ltd.

Fahrenheitstraße 13
28359 Bremen
GERMANY

www.campbellsci.de • info@campbellsci.de

Campbell Scientific do Brasil Ltda.

Rua Apinagés, n.br. 2018 — Perdizes
CEP: 01258-00 — São Paulo — SP
BRASIL

www.campbellsci.com.br • vendas@campbellsci.com.br

Campbell Scientific Spain, S. L.

Avda. Pompeu Fabra 7-9, local 1
08024 Barcelona
SPAIN

www.campbellsci.es • info@campbellsci.es

Please visit www.campbellsci.com to obtain contact information for your local US or international representative.