



**EU Declaration of Conformity**  
(in accordance with EN ISO/IEC 17050:2010)

**Product: HygroVUE™5**

**Manufacturer:**

Campbell Scientific, Ltd. | 80 Hathern Road, Shepshed, | Leics, UK, LE12 9GX |  
+44 1509 601141 Ph | +44 1509 601091 F | [www.campbellsci.co.uk](http://www.campbellsci.co.uk)

**This declaration of conformity is issued under the sole responsibility of the manufacturer for the following product(s):**

Model	Description
HygroVUE5	Temperature and Humidity Probe

**The object of the declaration described above is in conformity with the relevant Union harmonization legislation:**

**2014/30/EU**                      The Electromagnetic Compatibility Directive (EMC)  
**2011/65/EU**                      The Restriction of Hazardous Substances Directive (RoHS2)

**Standards referenced during compliance assessment:**

**EN 61326-1:2013**                      Electrical equipment for measurement, control and laboratory use –  
EMC requirements – for use in industrial locations.  
**EN 50581:2012**                      Technical documentation for the assessment of electrical and electronic  
products with respect to the restriction of hazardous substances.

**Signed for and on behalf of:**

**Place:** Campbell Scientific Ltd.  
**Date:** 23<sup>rd</sup> April 2019

Andrew Sandford  
Director of Research and Development  
Campbell Scientific Ltd.