

INSTRUCTION MANUAL



HUB-SDM5

7/04

Copyright (c) 2004
Campbell Scientific, Inc.

Warranty and Assistance

The **HUB-SDM5** is warranted by CAMPBELL SCIENTIFIC, INC. to be free from defects in materials and workmanship under normal use and service for twelve (12) months from date of shipment unless specified otherwise. Batteries have no warranty. CAMPBELL SCIENTIFIC, INC.'s obligation under this warranty is limited to repairing or replacing (at CAMPBELL SCIENTIFIC, INC.'s option) defective products. The customer shall assume all costs of removing, reinstalling, and shipping defective products to CAMPBELL SCIENTIFIC, INC. CAMPBELL SCIENTIFIC, INC. will return such products by surface carrier prepaid. This warranty shall not apply to any CAMPBELL SCIENTIFIC, INC. products which have been subjected to modification, misuse, neglect, accidents of nature, or shipping damage. This warranty is in lieu of all other warranties, expressed or implied, including warranties of merchantability or fitness for a particular purpose. CAMPBELL SCIENTIFIC, INC. is not liable for special, indirect, incidental, or consequential damages.

Products may not be returned without prior authorization. The following contact information is for US and International customers residing in countries served by Campbell Scientific, Inc. directly. Affiliate companies handle repairs for customers within their territories. Please visit www.campbellsci.com to determine which Campbell Scientific company serves your country. To obtain a Returned Materials Authorization (RMA), contact CAMPBELL SCIENTIFIC, INC., phone (435) 753-2342. After an applications engineer determines the nature of the problem, an RMA number will be issued. Please write this number clearly on the outside of the shipping container. CAMPBELL SCIENTIFIC's shipping address is:

CAMPBELL SCIENTIFIC, INC.

RMA#_____

815 West 1800 North

Logan, Utah 84321-1784

CAMPBELL SCIENTIFIC, INC. does not accept collect calls.

HUB-SDM5 Table of Contents

PDF viewers note: These page numbers refer to the printed version of this document. Use the Adobe Acrobat® bookmarks tab for links to specific sections.

1. Introduction	1
2. HUB-SDM5 Specifications	1
3. General	1
4. Installation and Wiring	1
4.1 Wiring SDM Devices	2
4.2 Wiring SDI-12 Devices	3
5. Generic Junction Box	4
6. Maintenance	4

HUB-SDM5

1. Introduction

The HUB-SDM5 is a junction box and is used to connect multiple Synchronous Device for Measurement (SDM) or SDI-12 devices/peripherals to a single Campbell Scientific, Inc. datalogger. The HUB-SDM5 consists of an environmental enclosure with six weather resistant bulkhead fittings and five terminal strips.

2. HUB-SDM5 Specifications

Terminal Strips:

Three spring-loaded guillotine connections per terminal; two terminals in common through a jumper creates five banks

Operating Temperature:

-55°C to +85°C

Operating Humidity:

Noncondensing 0% – 95%

Dimensions:

Length: 17.8 cm (7.0")
Width: 19.7 cm (7.75")
Depth: 13.3 cm (5.25")

Weight:

0.82 kg (1.8 lbs)

3. General

Inside the HUB-SDM5 enclosure are five pairs of terminal strips. These terminal strips are mounted on a DIN rail and an orange end cap separates each pair. Each pair of terminal strips is connected to each other by means of a jumper. The terminal strips are used to connect up to six wires to each other. The wires are secured to the terminal by a spring-loaded guillotine.

4. Installation and Wiring

A U-bolt is used to mount the HUB-SDM5 to a 1.25 inch NPT pipe.

The spring-loaded guillotines are opened with a small flat blade (2.5 X 50 mm) screwdriver (p/n 8125). Orient the screwdriver blade so that the blade edge is perpendicular to the orange separators. Gently insert the small flat blade

screwdriver into the square slot located next to the circular wire port until it bottoms out. For the cluster of four wire ports located on the left hand side of the HUB-SDM5, the guillotine release slot is located to the right of the wire port. For the wire ports on the right hand side of the HUB-SDM5, the guillotine release slot is located to the left of the wire port. Once the screwdriver has been completely inserted and has bottomed out, firmly press the screwdriver down until the guillotine snaps open. Insert and hold the wire as the screwdriver is removed.

4.1 Wiring SDM Devices

Before connecting the SDM signal cable to the terminals, strip the wire insulation back to expose ¼ inch of bare wire. Feed the cable through the bulkhead fitting. Neatly separate the five wires into the following order, SDM data, SDM clock, SDM enable, SDM ground, and SDM shield.




FIGURE 1. HUB-SDM5 connecting four SDM devices to a datalogger.




FIGURE 2. HUB-SDM5 terminal strip connecting four SDM devices to a datalogger.

4.2 Wiring SDI-12 Devices

Before connecting the SDI-12 signal/power cable to the terminals, strip the wire insulation back to expose ¼ inch of bare wire. Feed the cable through the bulkhead fitting. Neatly separate the four wires into the following order, SDI-12 data, SDI-12 power, SDI-12 ground, SDI-12 shield.




FIGURE 3. HUB-SDM5 connecting three SDI-12 devices to a datalogger.




FIGURE 4. HUB-SDM5 terminal strip with SDI-12 devices connected to a datalogger.

5. Generic Junction Box

The HUB-SDM5 can be used as a generic junction box. In its default configuration, six, five conductor cables can be connected to each other. Three, ten conductor cables can be connected to each other by removing the jumper between the terminal strips. Use a small screwdriver to pry the jumper out.

6. Maintenance

There are no user-serviceable parts in the HUB-SDM5. Make sure that all unused bulkhead fittings are sealed with a plug and that the lid is tightly fastened.

Campbell Scientific Companies

Campbell Scientific, Inc. (CSI)

815 West 1800 North
Logan, Utah 84321
UNITED STATES
www.campbellsci.com
info@campbellsci.com

Campbell Scientific Africa Pty. Ltd. (CSAf)

PO Box 2450
Somerset West 7129
SOUTH AFRICA
www.csafrica.co.za
sales@csafrica.co.za

Campbell Scientific Australia Pty. Ltd. (CSA)

PO Box 444
Thuringowa Central
QLD 4812 AUSTRALIA
www.campbellsci.com.au
info@campbellsci.com.au

Campbell Scientific do Brazil Ltda. (CSB)

Rua Luisa Crapsi Orsi, 15 Butantã
CEP: 005543-000 São Paulo SP BRAZIL
www.campbellsci.com.br
suporte@campbellsci.com.br

Campbell Scientific Canada Corp. (CSC)

11564 - 149th Street NW
Edmonton, Alberta T5M 1W7
CANADA
www.campbellsci.ca
dataloggers@campbellsci.ca

Campbell Scientific Ltd. (CSL)

Campbell Park
80 Hathern Road
Shepshed, Loughborough LE12 9GX
UNITED KINGDOM
www.campbellsci.co.uk
sales@campbellsci.co.uk

Campbell Scientific Ltd. (France)

Miniparc du Verger - Bat. H
1, rue de Terre Neuve - Les Ulis
91967 COURTABOEUF CEDEX
FRANCE
www.campbellsci.fr
campbell.scientific@wanadoo.fr

Campbell Scientific Spain, S. L.

Psg. Font 14, local 8
08013 Barcelona
SPAIN
www.campbellsci.es
info@campbellsci.es