

APPLICATION NOTE

Interfacing the AlphaGUARD Radon Monitor with Campbell Scientific's CR1000 Datalogger



Interfacing the AlphaGUARD Radon Monitor with Campbell Scientific's CR1000 Datalogger

This application note describes interfacing a CR1000 datalogger with a “standard version” of the AlphaGUARD Professional Radon Monitor Model P30 (not the “overdrive” version). The efforts to interface the AlphaGUARD Radon Monitor with the CR1000 Datalogger, and the writing of this application note were accomplished in collaboration with personnel from DRI, the environmental research arm of the Nevada System of Higher Education¹.

The AlphaGUARD is manufactured by Saphymo GmbH (www.saphymo.de)². This document is intended as a supplement to the manufacturer’s brochure and the manuals titled: “User Manual Portable Radon Monitor “AlphaGUARD dated 12/98” and “AlphaGUARD Data Transfer – Specifications version 4.0 dated 7/97”.

1. Radon Detection Basics

Two techniques are available for measuring radon activity; the best technique to use depends on the application. The techniques are integrative, passive radon sampling and continuous active radon sampling. Integrative, passive radon sampling does not require electrical power, whereas continuous active radon sampling does.

Examples of sensors that can measure radon levels include alpha track detectors, activated charcoal adsorption devices, and the AlphaGUARD monitor. Alpha track detectors are integrative, passive radon sampling devices that do not require ac power. They contain a thin piece of plastic or film mounted in the detector. Radon flows into the detector through a filtered opening. As the radon inside the detector decays, the emitted alpha particles hit the film forming tracks on the film. The tracks are counted to determine radon concentration. These detectors are not very effective in measuring high concentrations of radon [Spe04].

Activated charcoal devices are passive sensors that do not require ac power. They consist of a canister that holds granular-activated carbon (i.e., activated charcoal). The charcoal absorbs radon that enters the canister via a screened opening. After a determined exposure period, the canister is sealed and the charcoal is analyzed using a HPGe gamma ray detector [Ger04]. Activated charcoal devices are preferable for short-term measurements (i.e., 1 to 7 days). They are not practical for locations with high humidity because the charcoal can become saturated [Ger04].

The AlphaGUARD radon monitor is a continuous active radon sampling sensor. It has an ionization chamber and uses an alpha spectroscopy to detect radon. The two common isotopes of radon (i.e., ²²²Rn and ²²⁰Rn) are identified through their respective energies from the alpha decays [Swa04]. The signal generated from the alpha detection is converted to a digital output. This output can be read by the AlphaGUARD, dataloggers, or a PC. Ideal applications for the AlphaGUARD include radon exhalation, radon soil-gas, indoor, outdoor, in mining environments and in-water storage radon measurements.

¹DRI Project: Ambient Radon Monitoring in Amargosa Valley, Nevada
DRI Personnel: Greg McMurdy (greg.mccurdy@dri.edu), Jeff Tappen (Jeffrey.Tappen@dri.edu)

²AlphaGUARD Professional Radon Monitor Model P30 is manufactured by Saphymo GmbH, Heerstraße 149, D-60488, Frankfurt Germany, www.saphymo.de. In the U.S., the AlphaGUARD is available from Rad Elec, Inc., 5716-A Industry Lane, Frederick, Maryland 21704, www.radelec.com; phone (800) 526-5482, (301) 694-0011; fax (301) 694-0013. Approximate price is 6575 EUR or US \$9,886.

2. AlphaGUARD Sensor Description

The AlphaGUARD is designed for long-term monitoring of radon gas concentration. It has a linear response from 2 to 2,000,000 Bq/m³ (0.05 to 50,000 pCi/l). See the next section and the AlphaGUARD brochure for more specifications.

The AlphaGUARD measures radon concentration using 3D alpha spectroscopy. There is a lag in the signal due to the time required for radon to diffuse into the ionization chamber. When exposed to a step change in radon concentration, the AlphaGUARD will reach 30% of the new value after 10 minutes, 70% after 20 minutes and 90% after 30 minutes.

The AlphaGUARD has an internal battery that will run it for approximately 10 days. The unit can be AC powered with an AC to 12 Vdc wall charger from Radio Shack or CSI pn15966. Alternatively the AlphaGUARD may be DC powered using CSI pn14291. CSI pn14291 has a mating barrel connector on one end and red and black stripped and tinned leads to be connected to 12 Vdc and ground on the other end.



Figure 1. Front view of the AlphaGUARD Professional Radon Monitor Model P30 (standard version).

3. Specifications

The best source of specifications on the AlphaGUARD is the manufacturer's brochure which may be found at www.genitron.de/download/ag_fb_gb_01_a4.pdf. A few specifications are listed below for your convenience. Campbell Scientific recommends you refer to the manufacturer's brochure to verify critical specifications.

| | |
|------------------------|---|
| Measurement Range: | 2 to 2,000,000 Bq/m ³ (0.05 to 50,000 pCi/liter) |
| Operating temperature: | +14° to +122°F (-10° to +50°C) |
| Operating pressure: | 700 to 1100 mbar |
| Operating RH: | 0 to 99% RH |
| Operating voltage: | 12 Vdc (or 110 Vac with converter to 12 Vdc) |
| Current drain: | unknown (unit will run on internal battery for 10 days) |

4. Installation

The AlphaGUARD must be switched on manually (On/Off switch). Using the “Menu”, “Change”, and “Accept” buttons and the display, set the AlphaGUARD measuring cycle to 10 minutes and the baud rate to 9600 baud. The datalogger will time stamp the data but if it is important that the display indicate the correct time, set the AlphaGUARD clock using the serial interface software “DataEXPERT” provided with the unit.



The AlphaGUARD user manual indicates the automatic background correction of the AlphaGUARD can only be executed at the beginning of a measuring series. Therefore with measurements lasting longer than six months, the set up of a new measuring series should be manually enforced, by turning the power switch OFF and then on again after every half year. Setting the system clock with DataEXPERT will also initiate a new measuring series.

5. Wiring

The following wiring diagram is built for the program in Section 6 which is where the CR1000's COM1 port (C1 & C2) will request and capture data from AlphaGUARD's RS-232 port. The color scheme indicated in the last column is specific to the two foot long “DTE” cable which is one of two cables that come in Campbell Scientific's model SC110 (or pn16721).

| TABLE 1: AlphaGUARD to CR1000 Wiring | | | | |
|--------------------------------------|------------------|-----------------|-----------------|---------------------|
| AlphaGUARD Function | AlphaGUARD 9-pin | CR1000 Terminal | CR1000 Function | CSI SC110 DTE color |
| TxD | 2 | C2 | RxD | White |
| RxD | 3 | C1 | TxD | Brown |
| GND | 5 | GND | GND | Yellow |
| | | GND | Shell GND | Clear |



Figure 2. Above shows the AlphaGUARD connected to a CR1000 datalogger.

6. Example Program for the CR1000



This CR1000 program is for the “standard version” of the AlphaGUARD, not the “overdrive” version. It was tested on an AlphaGUARD with serial number 1635 EF.

```
'CR1000 Series Datalogger
'
' Program for Alphaguard Radon detector
Public fpval
Public sign
Public mantissa
Public exp1
Public exponent

Public Radon,RadonErr,AirTemp,RelHum,BaroPress
Public Akkustd,KBeqstd,Totalstd
Public ASerial as string * 6,Aversion as string * 4

Public PTemp, batt_volt
' Public HexV(80) as String *2 ': for debugging
DataTable (Test,1,-1)
  DataInterval (0,15,Sec,10)
  Minimum (1,batt_volt,FP2,0,False)
  Sample (1,PTemp,FP2)
EndTable

Public Str3 As String * 80
Public CurrDate as String * 12
Public DataDate as String * 12
Public Atemp, Arh
Public start As Long,i, rtnbyte(2), array(129) As Long
Alias array(8) = year
Alias array(9) = month
Alias array(10)=days
Alias array(11)=hrs
Alias array(12)=minutes
Alias array(13)=seconds
Alias array(14)=mesyear
Alias array(15)=mesmonth
Alias array(16)=mesday
Alias array(17)=meshr
Alias array(18)=mesmin
Alias array(19)=messec
Alias array(20)=notused
Alias array(28)=cycletime
Alias array(35)=DataQA
Alias array(36)=sysQA
```

Function FLP2 (mant,expo)

```

sign = IIF(expo AND &H80,1.0,-1.0)
mantissa = mant/256.0
    exp1 = (expo AND &H7F) - 64
        exponent = PWR(2,exp1)
        FLP2 = (sign * mantissa) * exponent

```

EndFunction

Function FLP4 (mant1,mant2,mant3,expo)

```

sign = IIF(expo AND &H80,1.0,-1.0)
mantissa = (mant3 + ((mant2*256)) + ((mant1*65536)))/16777216
    exp1 = (expo AND &H7F) - 64
        exponent = PWR(2,exp1)
        FLP4 = (sign * mantissa) * exponent

```

EndFunction

BeginProg

```

'Open COM1 at 9600 baud
SerialOpen (Com1,9600,3,10,1000) ' format for binary, 8-bit, no parity.

```

```

Scan (10,Sec,0,0)

```

```

    PanelTemp (PTemp,250)
    Battery (Batt_volt)

```

```

If TimeIntoInterval (0,1,Min)

```

```

    start = 1
endif

```

```

If start Then

```

```

' Output string to query Alphaguard
SerialFlush(com1) ' Clean out buffers
Str3 = CHR(2)
SerialOutBlock (Com1,Str3,1)
Delay (0,100,mSec)
str3= CHR(3) + CHR(83) + CHR(3) + CHR(91)
SerialOutBlock (Com1,Str3,4)

```

```

' Wait for return of data from Alphaguard

```

```

Delay (0,4,Sec)
i=1
rtnbyte(2)=SerialInChk(com1)

```

```

While rtnbyte(2)

```

```

    i=i+1
    SerialInBlock(Com1,array(1),1)

```

```

    MoveBytes (array(i),3,array(1),0,1)

```

```
' following line useful for debugging
'   hexV(i) = Hex(array(i))
   rtnbyte(2)=SerialInChk(com1)
Wend

start=0
CurrDate = right("0" & Hex(year),2) & right("0" & Hex(month),2) & right("0" & Hex(days),2)
CurrDate = CurrDate & right("0" & Hex(hrs),2) & right("0" & Hex(minutes),2) & right("0" & Hex(seconds),2)
DataDate = right("0" & Hex(mesyear),2) & right("0" & Hex(mesmonth),2) & right("0" & Hex(mesday),2)
DataDate = DataDate & right("0" & Hex(meshr),2) & right("0" & Hex(mesmin),2) & right("0" & Hex(messec),2)

Radon = FLP2(array(32),array(31))
RadonErr = FLP2(array(34),array(33))
AirTemp = FLP2(array(40),array(39))
RelHum = FLP2(array(46),array(45))

BaroPress = FLP4(array(42),array(43),array(44),array(41))
Akkustd = FLP4(array(48),array(49),array(50),array(47))
KBeqstd = FLP4(array(52),array(53),array(54),array(51))
Totalstd = FLP4(array(56),array(57),array(58),array(55))

ASerial = CHR(array(60)) & CHR(array(61)) & CHR(array(62)) & CHR(array(63)) & CHR(array(64)) & CHR(array(65))
AVersion = CHR(array(66)) & CHR(array(67)) & CHR(array(68)) & CHR(array(69))

EndIf

NextScan
EndProg
```

7. Maintenance and Troubleshooting

See AlphaGUARD User Manual.

8. References

- [Ger04] Gervino, G., Bonetti, R., Cigolini, C., Marino, C., Prati, P., Pruiti, L. (2004). Environmental radon monitoring: comparing drawbacks and performances of charcoal canisters, alpha-track and E-PERM detectors. *Nuclear Instruments and Methods in Physics Research A*, Vol. 518, pp 452 – 455.
- [Spe04] Speelman, W., (2004). Master of Science full thesis, Department of Physics, University of the Western Cape, Australia. Modeling and Measurement Of Radon Diffusion Through Soil for Application On Mine Tailings Dams, pp. 36 - 63. http://etd.uwc.ac.za/userfiles/modules/etd/docs/etd_init_9490_1178282920.pdf
- [Swa04] Swakon, J., Kozak, K., Paszkowski, M., Gradzinski, R., Łoskiewicz, J., Mazur, J., Janik, M., Bogacz, J., Horwacik, T., Olko, P. (2004). Radon concentration in soil gas around local disjunctive tectonic zones in the Krakow area. *Journal of Environmental Radioactivity*, Vol. 78, Issue 2, pp 137 – 149.