



EU Declaration of Conformity
(in accordance with EN ISO/IEC 17050:2010)

Product:

CPEC306 (with CR6-RF422)
CPEC310 (with CR6-RF422)

Manufacturer:

Campbell Scientific, Inc. | 815 W 1800 N | Logan, UT 84321-1784 | 435.227.9000 | 435.227.9001 F
www.campbellsci.com

This declaration of conformity is issued under the sole responsibility of the manufacturer for the following product(s):

Model	Description
CPEC306-RF422	Expandable Closed Path Eddy Covariance System w/EC155 & Pump Module, with CR6-RF422
CPEC310-RF422	Expandable Closed Path Eddy Covariance System w/EC155,Pump Module, & 3-Valve Module, with CR6-RF422

The object of the declaration described above is in conformity with the relevant Union harmonization legislation:

2014/53/EU Radio Equipment Directive
2011/65/EU The Restriction of Hazardous Substances Directive (RoHS2)

Standards referenced during compliance assessment:

EMC	EN 301 489-1 V1.9.2 (2011-09) in accordance with the specific requirements of EN 301 489-3 V1.4.1 (2002-08), EN 61326-1:2013
Safety	EN 61010-1:2010, EN 62311:2008
Spectrum	EN 300 220-2 V2.3.1 (2010-02)
RoHS2	EN 50581:2012

Signed for and on behalf of:
Place: Campbell Scientific Inc.
Date: 15th February 2018

Karsten Sorensen
Quality Assurance Manager
Campbell Scientific Inc.