



**EU Declaration of Conformity**  
(in accordance with EN ISO/IEC 17050:2010)

**Product: CS241DM**

**Manufacturer:**

Campbell Scientific, Inc. | 815 W 1800 N | Logan, UT 84321-1784 | 435.227.9000 | 435.227.9001 F  
[www.campbellsci.com](http://www.campbellsci.com)

**This declaration of conformity is issued under the sole responsibility of the manufacturer for the following product(s):**

Model	Description
CS241DM	PT-1000 Class A, Precision Back-of-Module Temperature Sensor with Digital Modbus RS-485 Output

**The object of the declaration described above is in conformity with the relevant Union harmonization legislation:**

- 2014/30/EU** The Electromagnetic Compatibility Directive (EMC)
- 2011/65/EU** The Restriction of Hazardous Substances Directive (RoHS2)
- 2015/863/EU** Amending Annex II of RoHS Directive 2011/65/EU: Phthalates Restrictions

**Standards referenced during compliance assessment:**

**EN 61326-2-3:2013** Electrical equipment for measurement, control and laboratory use – EMC requirements; Criteria for transducers with integrated or remote signal conditioning.

**EN 50581:2012** Technical documentation for the assessment of electrical and electronic products with respect to the restriction of hazardous substances.

**Signed for and on behalf of:**

**Place:** Campbell Scientific Inc.

**Date:** May. 4, 2021

Name: Gary Travis Roberts

Title/Position: Manager, Product Test and Compliance

Tel: (435) 227-9623

E-mail: [gtroberts@campbellsci.com](mailto:gtroberts@campbellsci.com)