



Declaration of Conformity
(In accordance with UK government legislation)

Product: CR300-RF422

Manufacturer:

Campbell Scientific, Inc. | 815 W 1800 N | Logan, UT 84321-1784 | 435.227.9000 | 435.227.9001 F
www.campbellsci.com

This declaration of conformity is issued under the sole responsibility of the manufacturer for the following product(s):

Model	Description
CR300-RF422	CR300 Datalogger (-40 to +70C) with 868 MHz SRD860 Radio

The object of the declaration described above is in conformity with the relevant UK Statutory Instruments and their amendments:

- 2017 No. 1206** The Radio Equipment Regulations 2017
- 2012 No. 3032** The Restriction of the Use of Hazardous Substances in Electrical and Electronic Equipment Regulations 2012

Standards referenced during compliance assessment:

EMC	EN 301 489-1 V1.9.2 (2011-09) in accordance with the specific requirements of EN 301 489-3 V1.4.1 (2002-08), BS EN 61326-1:2013
Safety	BS EN 61010-1:2010, BS EN 62311:2008
Spectrum	EN 328 220-2 V1.7.1 (2006-10)
RoHS	BS EN 63000:2018

Signed for and on behalf of:

Place: Campbell Scientific Inc.

Date: October 13, 2022

Name: Gary Travis Roberts

Title/Position: Manager, Product Test, and Compliance

Tel: (435) 227-9623

E-mail: gtroberts@campbellsci.com