

UK Declaration of Conformity

(In accordance with UK government legislation)

Product: IRGASON

This declaration of conformity is issued under the sole responsibility of the manufacturer (Campbell Scientific, Inc.) for the following product(s):

Model	Description
IRGASON	Integrated CO2 and H2O Open-Path Gas Analyzer and 3D Sonic Anemometer

The object of the declaration described above is in conformity with the relevant Union harmonization legislature:

2016 No. 1091	The Electromagnetic Compatibility Regulations 2016
2012 No. 3032	The Restriction of the Use of Hazardous Substances in Electrical and Electronic Regulations 2012

Standards referenced during compliance assessment:

BS EN 61326-1:2013	Electrical equipment for measurement, control, and laboratory use – EMC requirements – for use in industrial locations.
BS EN 63000:2018	Technical documentation for the assessment of electrical and electronic products with respect to the restriction of hazardous substances.

Signed for and on behalf of:



Gary Travis Roberts

Manager, Product Test, and Compliance

Campbell Scientific, Inc. | Public

www.campbellsci.com | ph: (435) 227-9000 fax: (435) 227-9001 | 815 W 1800 N, Logan, UT 84321

Australia | Brazil | Canada | China | Costa Rica | France | Germany | India | South Africa | Spain | Thailand | UK | USA