

## **UK Declaration of Conformity**

(In accordance with UK government legislation)

## **Product: TDR200**

This declaration of conformity is issued under the sole responsibility of the manufacturer (Campbell Scientific, Inc.) for the following product(s):

| Model  | Description               |
|--------|---------------------------|
| TDR200 | Time-Domain Reflectometer |

## The object of the declaration described above is in conformity with the relevant Union harmonization legislature:

| <b>2016 No. 1091</b> The Electromagnetic Compatibility Regulations 2016 |  |
|---|--|
|---|--|

2012 No. 3032 The Restriction of the Use of Hazardous Substances in Electrical and

Electronic Regulations 2012

## Standards referenced during compliance assessment:

BS EN 61326-1:2013 Electrical equipment for measurement, control, and laboratory use – EMC

requirements - for use in industrial locations.

**BS EN 63000:2018** Technical documentation for the assessment of electrical and electronic

products with respect to the restriction of hazardous substances.

Signed for and on behalf of:

**Gary Travis Roberts** 

Manager, Product Test, and Compliance