

INSTRUCTION MANUAL

Model CS205 10-Hour Fuel Temperature Stick

Revision: 5/08



Copyright © 1998-2008
Campbell Scientific, Inc.

Warranty and Assistance

The **MODEL CS205 10-HOUR FUEL TEMPERATURE STICK** is warranted by CAMPBELL SCIENTIFIC, INC. to be free from defects in materials and workmanship under normal use and service for twelve (12) months from date of shipment unless specified otherwise. Batteries have no warranty. CAMPBELL SCIENTIFIC, INC.'s obligation under this warranty is limited to repairing or replacing (at CAMPBELL SCIENTIFIC, INC.'s option) defective products. The customer shall assume all costs of removing, reinstalling, and shipping defective products to CAMPBELL SCIENTIFIC, INC. CAMPBELL SCIENTIFIC, INC. will return such products by surface carrier prepaid. This warranty shall not apply to any CAMPBELL SCIENTIFIC, INC. products which have been subjected to modification, misuse, neglect, accidents of nature, or shipping damage. This warranty is in lieu of all other warranties, expressed or implied, including warranties of merchantability or fitness for a particular purpose. CAMPBELL SCIENTIFIC, INC. is not liable for special, indirect, incidental, or consequential damages.

Products may not be returned without prior authorization. The following contact information is for US and International customers residing in countries served by Campbell Scientific, Inc. directly. Affiliate companies handle repairs for customers within their territories. Please visit www.campbellsci.com to determine which Campbell Scientific company serves your country. To obtain a Returned Materials Authorization (RMA), contact CAMPBELL SCIENTIFIC, INC., phone (435) 753-2342. After an applications engineer determines the nature of the problem, an RMA number will be issued. Please write this number clearly on the outside of the shipping container. CAMPBELL SCIENTIFIC's shipping address is:

CAMPBELL SCIENTIFIC, INC.

RMA# _____

815 West 1800 North

Logan, Utah 84321-1784

CAMPBELL SCIENTIFIC, INC. does not accept collect calls.

CS205 Table of Contents

PDF viewers note: These page numbers refer to the printed version of this document. Use the Adobe Acrobat® bookmarks tab for links to specific sections.

- 1. General Description..... 1
- 2. Specifications 1
- 3. Installation..... 1
- 4. Wiring..... 2
- 5. Datalogger Programming..... 3
 - 5.1 CR1000 Programming..... 3
 - 5.2 CR10X Programming..... 4
- 6. Maintenance and Calibration..... 4

Figures

- 3-1. Exploded View of CS505 and CS205 Mounting 1
- 3-2. CS205 Mounts on the Mounting Stake with the CS505..... 2
- 6-1. 107 Thermistor Probe Schematic 4

Tables

- 4-1. Wiring for CS205 2

Model CS205 10-hour Fuel Temperature Stick

1. General Description

The CS205 Temperature Stick is a Ponderosa pine dowel 4.5 inches long and 0.5 inches in diameter. The dowel is fabricated to USFS specifications, to be used as a fuel moisture sensor. Campbell Scientific, Inc. has bored a hole and cut a split into one end of the dowel. A model 107 Temperature Probe is inserted into the CS205 to measure fuel temperature.

The CS205 Fuel Temperature Stick mounts on the 10974 Fuel Moisture/Temperature Mounting Stake. The CS205 is most commonly installed along with a CS505 Fuel Moisture Sensor but can be used alone.

2. Specifications

Fuel Temperature Stick	
Diameter	0.5 inches
Length	4.5 inches
Wood Type	Ponderosa Pine
Bore Hole	about 1/4 inch

3. Installation

Insert a 107 temperature probe into the CS205 Fuel Moisture Stick. The CS205 installs in the white compression fitting on the 10974 Fuel Moisture/Temperature Mounting Stake. Loosen the compression fitting so that the CS205 can be inserted. The CS205 inserts into the compression fitting so that compression is applied to the split end of the stick gripping the 107 probe.

Please refer to the 107 probe manual for a complete discussion of the 107 probe.




FIGURE 3-1. Exploded View of CS505 and CS205 Mounting




FIGURE 3-2. CS205 Mounts on the Mounting Stake with the CS505

4. Wiring

Connections to Campbell Scientific dataloggers are given in Table 4-1. Temperature is measured with one Single-Ended input channel and an Excitation channel.

TABLE 4-1. Wiring for CS205				
Color	Description	CR800 CR5000 CR3000 CR1000	CR510 CR500 CR10(X)	CR21X CR7 CR23X
Black	Excitation	Switched Excitation	Switched Excitation	Switched Excitation
Red	Temperature Signal	Single-Ended Input	Single-Ended Input	Single-Ended Input
Purple	Signal Ground	⏏	AG	⏏
Clear	Shield	⏏	G	⏏

Please refer to the 107 probe manual for a complete discussion of the 107 probe.

5. Datalogger Programming

This section is for users who write their own datalogger programs. A datalogger program to measure this sensor can be created using Campbell Scientific's Short Cut Program Builder software. You do not need to read this section to use Short Cut.

The Temp107 measurement instruction (P11), is used with dataloggers that are programmed with Edlog (e.g. CR10X, CR23X) to measure the 107 probe. P11 makes half bridge voltage measurement, and converts the measurement result to temperature using a fifth order polynomial. With a multiplier of 1 and an offset of 0, the output is temperature in degrees C. With a multiplier of 1.8 and an offset of 32, the output is temperature in degrees F.

The Therm107 measurement instruction is used with dataloggers that are programmed with CRBasic (e.g. CR1000) to measure the 107 probe. Therm107 makes a half bridge voltage measurement, and converts the measurement result to temperature using the Steinhart-Hart equation. With a multiplier of 1 and an offset of 0, the output is temperature in degrees C. With a multiplier of 1.8 and an offset of 32, the output is temperature in degrees F.

5.1 CR1000 Programming

```
'CR1000
'This example program measures a single 107 Thermistor probe
'every 10 seconds and stores the average temperature every 60 minutes.

'Declare the variables for the temperature measurement
Public T107_C

'Define a data table for 60 minute averages:
DataTable(Table1,True,-1)
    DataInterval(0,60,Min,0)
    Average(1,T107_C,IEEE4,0)
EndTable

BeginProg
    Scan(10,Sec,1,0)
        'Measure the temperature
        Therm107(T107_C,1,1,Vx1,0,_60Hz,1.0,0.0)
        'Call Data Table
        CallTable(Table1)
    NextScan
EndProg
```

5.2 CR10X Programming

```

;{CR10X}
*Table 1 Program
01: 10      Execution Interval (seconds)

1: Temp (107) (P11)
1: 1      Repts
2: 1      SE Channel
3: 21     Excite all reps w/E1, 60Hz, 10ms delay
4: 1      Loc [ T107_C ]
5: 1.0    Multiplier
6: 0.0    Offset

3: If time is (P92)
1: 0      Minutes (Seconds --) into a
2: 60     Interval (same units as above)
3: 10     Set Output Flag High (Flag 0)

4: Set Active Storage Area (P80)
1: 1      Final Storage Area 1
2: 101    Array ID

5: Real Time (P77)
1: 1220   Year,Day,Hour/Minute (midnight = 2400)

6: Average (P71)
1: 1      Repts
2: 1      Loc [ T107_C ]

```

6. Maintenance and Calibration

The 107 Probe requires minimal maintenance. Please refer to the 107 manual for more detail.

The CS205 Fuel Moisture Stick should be changed annually or more frequently as required. The wood should visually appear fresh and new not gray or discolored.




FIGURE 6-1. 107 Thermistor Probe Schematic

Campbell Scientific Companies

Campbell Scientific, Inc. (CSI)

815 West 1800 North
Logan, Utah 84321
UNITED STATES
www.campbellsci.com
info@campbellsci.com

Campbell Scientific Africa Pty. Ltd. (CSAf)

PO Box 2450
Somerset West 7129
SOUTH AFRICA
www.csafrica.co.za
cleroux@csafrica.co.za

Campbell Scientific Australia Pty. Ltd. (CSA)

PO Box 444
Thuringowa Central
QLD 4812 AUSTRALIA
www.campbellsci.com.au
info@campbellsci.com.au

Campbell Scientific do Brazil Ltda. (CSB)

Rua Luisa Crapsi Orsi, 15 Butantã
CEP: 005543-000 São Paulo SP BRAZIL
www.campbellsci.com.br
suporte@campbellsci.com.br

Campbell Scientific Canada Corp. (CSC)

11564 - 149th Street NW
Edmonton, Alberta T5M 1W7
CANADA
www.campbellsci.ca
dataloggers@campbellsci.ca

Campbell Scientific Ltd. (CSL)

Campbell Park
80 Hathern Road
Shepshed, Loughborough LE12 9GX
UNITED KINGDOM
www.campbellsci.co.uk
sales@campbellsci.co.uk

Campbell Scientific Ltd. (France)

Miniparc du Verger - Bat. H
1, rue de Terre Neuve - Les Ulis
91967 COURTABOEUF CEDEX
FRANCE
www.campbellsci.fr
info@campbellsci.fr

Campbell Scientific Spain, S. L.

Psg. Font 14, local 8
08013 Barcelona
SPAIN
www.campbellsci.es
info@campbellsci.es